

杭州奥泰生物技术股份有限公司 Hangzhou AllTest Biotech Co.,Ltd	文件号 Document No.: ZTC-QC-005-R-003
传染病、心肌、肿瘤类 COA The Infectious Disease、Cardiology、Tumor COA	生效日期 Effective Date: 2018 年 07 月 02 日

Certificate of Analysis

Product Name:SARS-CoV-2 and Influenza A+B Antigen Combo Rapid Test(Nasal Swab)

Catalog No.: ISIN-525H

Batch No.: ATSIN22090006

Quantity:100000PCS

Expiry Date:2024-08

Date of Sampling:2022-09-09

Date of Analysis:2022-09-09

Other information:/

QC Item		QC Criterion	QC Result	Conclusion
Physical	Appearance	Good	Good	Pass
NCP	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
FLU	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass

Others	/
---------------	---

Final QC Conclusion:	This batch of product met the QC Criteria.
-----------------------------	--

QC supervisor: Freeman.zheng

Date: 2022.09.09

